



PRECAUTION TIPS



Check the service indicator light



Check for any wear and tear on the AED



Ensure that AED pads and accessories are not expired



Consider adding additional AEDs in high traffic areas such as hallways, main building entrances, gymnasiums, cafeterias, auditoriums, and nurse's office

Verify that the AED batteries are working



